

**AMS 2431/1D & 2431/2D section 3.9.1 current wording**

The size of shot, specified in 3.8, shall be determined by using a 100-gram (approximately) sample and screening as follows: The required standard testing sieves in accordance with ASTM E 11 shall be nested in ascending order with a pan on the bottom. The 100-gram sample shall be poured onto the top sieve and the nested sieves shall be placed in a rotating and tapping type of shaking machine. **The rotating speed shall be 275 to 295 rpm and the tapping speed shall be 145 to 160 taps per minute.** Shaking and tapping shall be continued for 5 minutes  $\pm$  5 seconds for sieves 30 mesh and coarser and 10 minutes  $\pm$  5 seconds for sieves finer than 30 mesh. After shaking, the percentage of shot on each screen shall be determined by weighing the shot retained on each screen.

**Proposed wording**

The size of shot, specified in 3.8, shall be determined by using a 100-gram (approximately) sample and screening as follows: The required standard testing sieves in accordance with ASTM E 11 shall be nested in ascending order with a pan on the bottom. The 100-gram sample shall be poured onto the top sieve and the nested sieves shall be placed in a rotating and tapping type of shaking machine. **The rotating speed shall be 270 to 300 rpm and the tapping speed shall be 140 to 160 taps per minute.** Shaking and tapping shall be continued for 5 minutes  $\pm$  5 seconds for sieves 30 mesh and coarser and 10 minutes  $\pm$  5 seconds for sieves finer than 30 mesh. After shaking, the percentage of shot on each screen shall be determined by weighing the shot retained on each screen.

**Revise AMS 2431 section 3.3.4 to add:**

The following items are **required** to produce an accurate sample for particle analysis. **See Table I for applicability.**

- 16-1 Sample Reducer (to reduce bulk material i.e. 50 pound bag to 3.125 pounds in a single pass)
- Sample Splitter or Jones type riffle ( This is used to split the sample in half while maintaining the particle size distribution of the original sample)
- Sieve shaker (rotating and tapping type with digital timer with accuracy of +/-5 seconds)
- Sieves set in accordance with ASTM E-11
- Digital Balance having a capacity of at least 150 g and a sensitivity of 0.01 g.
- Stereoscopic Microscope capable of at least 30x magnification
- Standard Micrometer

<b>Equipment</b>	16-1 Reducer	Sample Splitter	Sieve Shaker	Sieve Set	Digital Balance	Stereoscopic Microscope	Standard Micrometer
<b>Specification</b>							
AMS 2431/1	YES	YES	YES	YES	YES	YES	NO
AMS 2431/2	YES	YES	YES	YES	YES	YES	NO
AMS 2431/3	YES	YES	NO	NO	YES	YES	NO
AMS 2431/4	YES	YES	NO	NO	YES	YES	NO
AMS 2431/5	YES	YES	NO	NO	YES	NO	YES
AMS 2431/6	YES	YES	YES	YES	YES	YES	NO
AMS 2431/7	YES	YES	YES	YES	YES	YES	NO
AMS 2431/8	YES	YES	NO	NO	YES	YES	NO

TABLE I

## Rational

AMS 2431B

**Section 3.3.4 Size shall be in accordance with ASTM E 11 , ASTM B 214, and the applicable detail specification.**

- ASTM B 214 calls for The rotating speed to be 270 to 300 rpm and the tapping speed to be 140 to 160 taps per minute.

By revising AMS 2431, /1 & /2 to be in accordance with ASTM B 214 we remove the dual dimensioning issue between the two. It also resolves the issue the Tyler Ro-Tap. Potentially resolves European issues regarding the differences in electrical current.

Specification	Section	RPM	Taps per Minute
ASTM B214-07	5.2	270-300	140-160
AMS 2431B	3.3.4 (ASTM B214)	270-300	140-160
AMS 2431/1D	3.9.1	275-295	145-160
AMS 2431/2D	3.9.1	275-295	145-160
SAE J444	3.1.1.1 / 3.1.1.2	275-295	145-160
AC7117	6.1.3 (SAE J444)	275-295	145-160
Tyler Ro-Tap		270-290	145-155

Table II

Table II shows that the base requirement is 270-300 RPM / TAPS per minute of 140-160 the industry standard sieve tester is the Ro-Tap which per the manufacture operates at 270-290 RPM and 145-155 Taps per minute.

SAE J444 and Nadcap AC7117 will also need to be revised.

**AMS 2431B** does not clearly identify the all necessary accessory equipment needed to produce and process an accurate sample of peening media.